

User Manual

Mandala Dynamic Scan




Please read the manual before using the tools.

Version 4.3

Table of Contents

1	General Notes.....	2
1.1	Product Feature and Intend Use	2
1.2	Warning	2
1.3	Warranty	2
1.4	Precautions.....	2
1.5	Maintenance.....	3
1.6	Product Include.....	4
2	Technical Data	5
2.1	Specifications	5
2.2	Overview and Function.....	6
3	Installation and Operation Instruction	7
3.1	Setup	7
3.2	Charging Setup	9
4	Fault Diagnosis.....	11



1 General Notes

1.1 Product Feature and Intend Use

Thank you for using Light Mandalas Product. Mandala Dynamic Scan has been in good condition before leaving factory.

Mandala Dynamic Scan is a tablet-based Bio resonance evaluation tool detecting the degrees of deviation between the user's Bio-matrix Field and libraries of resonant signatures identifying the most beneficial homeopathic remedies, Essential oils, Flower essences, crystals, emotional, mental and energetic (psycho-spiritual) issues, allergens, toxins and detects the body's response. Then produces a report.

Mandala Dynamic Scan can be used with other Light Mandalas devices to scan the best remedies or healing presets for the user at that present moment. It also can pair with Mandala Light Alchemy for biofeedback scan by Bluetooth. In this way, users can observe their real time body coherence status during meditation, yoga, and exercise.

1.2 Warning

This product is in perfect condition before delivery, please check all accessory when unpacking the unit.

Pay close attention to each safety REMARK and WARNING statement in the manual. This manual contains important operation information.

1.3 Warranty

Each single product passes through a quality process. Nevertheless, if failures occur, contact the customer service at once. The product is under 1 year warranty effectively according to the delivery date stated on the warranty card.

Warranty does not apply to damages, which result from misuse or neglect. The warranty also expires if you dismantle the product or failed to provide the warranty

card. The manufacturer is not liable for consequential damage or in case of a non-intended use of the product.

If a failure occurs during the warranty period, the product will be replaced or repaired. The manufacturer reserves the right to exchange components of the product instead of repairing it. If the failure results from misuse or neglect, the user must pay for all cost that may occur. The freight cost and other costs will be depending on which Term & Conditions criteria the product falls into. **Import tax and Customs clearance duty cost is not included in the term of warranty.** In that case, you may ask for an estimated cost beforehand.

1.4 Precautions

1. Please check Mandala Dynamic Scan for damage during shipping. Do not use the device if any damage is discovered. If the device is damaged, please contact the dealer or Light Mandalas immediately
2. Please Confirm the product you have received are in correct voltage and frequency for your country.
3. Mandala Dynamic Scan needs to be charged for 2 hours prior to first use.
4. Do not disassemble or modify the device in any case.
5. Do not use the device outdoors when rains
6. Places the device in a well-ventilated area.
7. Keep dry and avoid excessive humidity, heat or dust.
8. Do not place the device closer than 2 feet from combustible object.

1.5 Maintenance

1. If the device has not been used for a long time, please charge for 2 hours before use.
2. Clean with soft cloth, humid tissue moistened with water or a water-based glass cleaner.
3. In order to guarantee the fixture works properly, it's essential to keep it clean.

1.6 Product Include

- | | |
|------------------------------|----|
| 1. Mandala Dynamic Scan | x1 |
| 2. Ear Clip Sensor | x1 |
| 3. USB Charging cable | x1 |
| 4. Charging Adapter | x1 |
| 5. Mandala Dynamics Scan Box | x1 |



2 Technical Data

2.1 Specifications

Power Supply:	Charger Adapter 100-240VAC, 50-60 Hz convert to 5V 2A DC power
Working Mode:	Thumb Sensor and Ear Clip Sensor
Case Color:	Black
Case Material:	Aluminum and Acrylic black
Size:	115 x 65 x 25 mm
Weight:	400 g



2.2 Overview and Function

Top View




- | | |
|--|--|
| <p>1. Thumb Sensor:</p> <p>2. Body Coherence Indicator:</p> <p>3. HRV Battery Indicator:</p> <p>4. Heart Rate Pulse Indicator:</p> | <p>For thumb scan use</p> <p>Display body coherence status
 Red: Mental/Emotional level of functioning.
 Blue: Heart, mind and soul beginning to connect.
 Green: Coherent Stage. Merge between heart, mind and soul</p> <p>Display HRV battery status</p> <p>Display pulse detection</p> |
|--|--|

Right Side View



Left Side View



- | | |
|---|---|
| <p>1. ON/OFF Button:</p> <p>2. USB Charging Connector:</p> <p>3. Ear Sensor Port:</p> <p>4. Bluetooth Battery Charging Indicator:</p> | <p>Turn ON/OFF the device. (press and hold for 3 seconds to turn ON or OFF)</p> <p>Plug in charging cable for charging use</p> <p>Plug in ear clip sensor to use ear sensor</p> <p>Display bluetooth battery charging status
 RED – Charging; GREEN – Fully Charge</p> |
|---|---|

3 Installation and Operation Instruction

3.1 Setup

The device can be used to scan with either **Ear Sensor** or **Thumb Sensor**.

Ear Sensor:


1. To use this sensor, plug in the Ear Sensor cable to the Ear Sensor Port.



2. Clip the ear sensor on either one of your earlaps.



We recommend attaching the lapel clip to your garment to offset movement.



Thumb Sensor:

To use this sensor, gently place your thumb on the Thumb Sensor.



The ear sensor should not be plugged in if using the thumb sensor to detect a pulse.



1. When sensor is set, press the on/off button and hold for 3 seconds to turn on.



The device detects body coherence by HRV (Heart Rate Variability) and it will turn off automatically to save battery when receive no pulse signal in a short period of time.

2. Ensure the **Heart Rate Pulse Indicator** is blinking blue, indicating your **Mandala Dynamic Scan** is reading your pulse.



If it doesn't blink blue, adjust the ear sensor or your thumb position over the Thumb Sensor

3. Your Mandala Dynamic Scan is ready for scan now, you may connect to **Mandala Star Application** and start to scan.

3.2 Charging Setup

There are 2 batteries in this device – **Sensor Battery** and **Bluetooth Battery**.

1. To charge Mandala Dynamic Scan, plug in the **Mini USB Charging Cable** to the device charging port and connect the other end to the provided **Charging Adapter** or powered up computer.



2. Connect the **Charging Adapter** to the power supply and turn it on.
3. The sensor battery status will be shown on the **Battery Indicator** when the device is charging. To confirm proper connection, you must see 5 blue lights on the Battery Indicator. The 5 blue lights indicate charge status and connectivity. When device reaches a higher charge, the lights will become solid (not blinking).


5 blue lights cycling up and down → charging

5 blue lights are solid (not blinking) → fully charged



Battery Status

4. The **Bluetooth Battery Charging Status** will be shown on the left side of the device when it's plugged into charging adapter.



RED – Charging GREEN – Fully Charged

5. The charging process takes about 2 hours if starting with a fully drained battery and provides over 6 hours of battery life.



In order to protect the battery life span, please don't overcharge the battery.

4 Fault Diagnosis

Failures	Remedies
<p>Device turn off by itself (Device is designed to turn off automatically if it can't detect any pulse in a short period of time)</p>	<p>a) check Heart Rate Pulse Indicator. Is it pulsing? If not, means the device can't detect your heart rate. This may be due to the following:</p> <ul style="list-style-type: none"> • device battery life is low. Please charge the device. • ear clip plug loose or thumb position is not correct. Remove the ear clip sensor and plugging back in again or replace your thumb position if you use Thumb Sensor. <p>b) Plug in the charger. If it can be used during charging, that means the device need to be charged.</p>
<p>Cannot connect to tablet</p>	<p>a) Check the application setting.</p> <p>b) Turn off the device, and turn off tablet Bluetooth then try to turn them on again.</p> <p>c) Plug in the charger. If it can be used during charging, that means the Bluetooth battery need to be charged.</p>

Disclaimer

All content found on the Light-mandalas.com Website, Facebook Page, YouTube, including: text, images, video, audio, or other formats are intended for informational, educational and research purposes only and neither Light Mandalas Co. Ltd. nor the publishers of this content take responsibility for possible health consequences of any person or persons reading or following the information in this educational content and is not intended to be a substitute for professional medical advice, diagnosis, treatment or cure. Always seek the advice of your physician or other qualified health professionals with any questions you may have regarding a medical condition.

“No claims other than stress detection and stress reduction may be made”

Light Mandalas products are to be used as a Heart Rate Variability Biofeedback system and Light, Color and Sound therapy, designed for stress detection and stress reduction only. The devices do not evaluate any issue other than stress. Stress can come from many sources; this system uses light and color and sound therapies to reduce stress. Remember, there has never been any reported side effects using light, color and sound therapy anywhere in the world and throughout time.

Notes on the practices of low-level laser therapy (LLLT) and Biofeedback

Low-level laser therapy (LLLT) is considered investigational by the FDA, and no medical claims of cures are permitted in the US; at this time however, low-level lasers are defined within the scope of complimentary medical practices such as physiotherapy, aesthetic, Traditional Chinese Medicine and acupuncture, etc.

The FDA does not regulate Bio resonance tools but they do regulate biofeedback tools and some Bio resonance tools are registered under this category.

*It is a legal requirement at this time that the user does not make any medical claims of diagnosis or cure.

Precaution

Light Mandalas products are working on energetic level. The standard benefit is energy clearance. During the process, your body and mind can be uncomfortable. The symptoms associated with overly rapid clearance of toxins. If those detox symptoms are too strong, you may need to stop using the products. You have to allow the clearance or detox to proceed and you can start to try the products afterwards. However, if the issues still remain the same, it may need to be completely discontinued for your own safety

Shall there be any inquiry, please contact us for further support.

Light Mandalas Co., Ltd. (Thailand Headquarters):

- **Light Mandalas Website:** www.light-mandalas.com
- **Official:** info@light-mandalas.com
- **Instagram:** [@mandalawave9](https://www.instagram.com/mandalawave9)
- **YouTube Channel:** Light Mandalas Official
- **Technical Support Email:** cservice@light-mandalas.com

Local agents contacts:

- **Japan:** japan@light-mandalas.com
- **Taiwan:** lovelight777888@gmail.com
- **Indonesia:** lightseed2020bali@gmail.com
- **West Malaysia:** lightworker@healingtech.org
- **East Malaysia:** malaysia@light-mandalas.com
- **Hungary:** horesnyibela@light-mandalas.com
- **Switzerland, Germany, Austria, and the Netherlands:** info@light-mandalas.ch
- **France, Spain:** france-spain@light-mandalas.com
- **USA:** usa@light-mandalas.com
- **New Zealand, Australia:** tachyonhealing.nz@gmail.com